Review Article

J. innov.appl. pharm. Sci, 11[3] 2026, 80-85

Journal of Innovations in Applied Pharmaceutical Science [JIAPS]

Content available at: www.saap.org.in ISSN: 2455-5177

# A COMPREHENSIVE REVIEW OF LIQUID CHROMATOGRAPHY-MASS SPECTROMETRY

**THIRUNAHARI PRATHIMA[1], KARINGALA SHOUREYA[2], VADAPALLI EKSHITHA[2], SUDHEER KUMAR KAMARAPU[1*]**

[1]*Sri Kakatiya Institute of Pharmaceutical Sciences, Unikicherla, Hanamkonda, Warangal - 506370, Telangana, India.*

[2]*School of Pharmacy, Anurag University, Venkatapur, Ghatkesar, Medchal– Malkajgiri, Hyderabad – 500088, Telangana, India.*

**Article History**

Received: 04-08-2026

Revised: 24-09-2026

Accepted: 16-10-2026

***Keywords:*** *Liquid chromatography-mass spectrometry, analytical chemistry, drug discovery, quadrupole, bioanalytical studies.*

***Corresponding Author**

Dr. Sudheer Kumar Kamarapu

**Abstract**

Liquid Chromatography–Mass Spectrometry (LC–MS) is an advanced analytical technique that combines the separation capability of liquid chromatography (LC) with the identification and detection capability of mass spectrometry (MS). It is widely used in pharmaceutical analysis, drug discovery, bioanalysis, metabolomics, proteomics, and clinical research. Liquid chromatographers choose liquid chromatography-mass spectrometry (LC-MS), which is rapidly evolving. The method known as liquid chromatography-mass spectrometry (LC-MS/MS) combines mass spectrometry with liquid chromatography, also known as HPLC. It is an analytical chemistry method that combines the mass analysis skills of mass spectrometry with the physical separation capabilities of liquid chromatography, also known as HPLC. (LC-MS/MS) is frequently used in labs for both qualitative and quantitative investigation of biological materials, drug compounds, and drug products.This review provides a quick overview of quantitative analysis techniques, which are among the most beneficial courses for scientific education. An overview of HPLC's applications and stationary phase is given. A brief discussion of the fundamental idea of mass spectrometry is given. This review's primary goal is to go over the most popular hyphenation method, liquid chromatography in combination with mass spectrometry [LC-MS]. The interface is crucial to the hyphenation process because it facilitates the transfer of eluent from liquid chromatography to a mass spectrometer.

This article is licensed under a Creative Commons Attribution-Non Commercial 4.0 International License.

Copyright © 2026 Author[s] retain the copyright of this article.

## INTRODUCTION

The combination of MS with chromatographic methods has consistently been sought after due to the sensitive and exceptionally specific qualities of MS in comparison to other chromatographic detectors. The integration of Gas Chromatography with MS (GC-MS) was accomplished in the 1950s, with commercial devices available starting in the 1970s. Affordable and dependable GC-MS systems are currently present in numerous clinical biochemistry labs and are essential in various fields where thorough analysis of complex mixtures and clear identification are necessary, such as testing urine samples for metabolic disorders or drugs [1,2]. The combination of MS with LC (LC-MS) was a natural progression, yet advancements in this field were constrained for many years because of the incompatibility of current MS ion sources with a steady liquid flow. Multiple interfaces were created, yet they were difficult to operate and inconsistent, resulting in minimal adoption by clinical laboratories. The development of the electrospray ion source by Fenn in the 1980s altered this situation. Producers quickly created devices featuring electrospray sources, significantly influencing protein and peptide biochemistry. Fenn received the Nobel Prize in 2002 alongside Koichi Tanaka, who created matrix assisted laser desorption ionization, a highly beneficial MS ionization method for studying biological molecules [2-4].

Liquid Chromatography–Mass Spectrometry (LC–MS) is a powerful analytical technique that combines the separation efficiency of liquid chromatography (LC) with the sensitive detection and identification capabilities of mass spectrometry (MS). It is widely used in pharmaceutical sciences, drug discovery, bioanalysis, clinical research, toxicology, proteomics, and metabolomics for the analysis of complex mixtures.Liquid chromatography separates the individual components of a sample based on their differential interactions with the stationary phase and mobile phase [5,6]. The separated compounds are subsequently introduced into the mass spectrometer, where they are converted into gas-phase ions and analyzed according to their mass-to-charge ratio (m/z). The resulting mass spectra provide valuable information for compound identification, molecular mass determination, and structural characterization. When tandem mass spectrometry (LC–MS/MS) is employed, additional selectivity and sensitivity can be achieved through the fragmentation of selected precursor ions into product ions [6-8].





Figure 01: Direct electron ionisation Liquid Chromatography-Mass Spectroscopy

LC–MS has become an essential tool in pharmaceutical analysis because it enables the detection and quantification of active pharmaceutical ingredients, impurities, degradation products, drug metabolites, and biological molecules, even at low concentrations. It is extensively applied in pharmacokinetic studies, bioavailability and bioequivalence investigations, therapeutic drug monitoring, drug metabolism studies, and the development of stability-indicating analytical methods. Compared with conventional chromatographic techniques using ultraviolet detection, LC–MS can provide greater molecular specificity and additional structural information for suitable analytes.

Despite its advantages, LC–MS has certain limitations, including high instrument and maintenance costs, the requirement for skilled personnel, and potential matrix effects such as ion suppression or enhancement. Nevertheless, continuous advancements in ionization techniques, mass analyzers, and data-processing software have significantly expanded its analytical capabilities. Therefore, LC–MS remains an indispensable technique in modern pharmaceutical research, quality control, and biomedical analysis.

The interfacing mechanism is located within a shared EI source, similar to those present in any GC-MS system. The liquid phase from a nano HPLC column enters through the capillary column port, where the connection tubing and nebulizer are initially inserted and sealed to avoid vacuum loss. The system relies on creating an aerosol under high-vacuum conditions, which is then subjected to rapid desolvation of droplets and subsequent vaporization of the solute before ionization occurs. The process concludes swiftly and thoroughly, lowering the risk of thermal decomposition as illustrated in Figure 1, which presents a diagram of the interface. The micro-nebulizer is at the center of the interface. The tip of the nebulizer extends into the ion source, ensuring that the spray expansion remains fully enclosed within the ion volume [8-12].




Figure 02: Liquid Chromatography-Mass Spectroscopy

The eluate appears as a liquid phase at a flow rate of 300-500 nL/min, and any early in-tube solvent evaporation is avoided by effective thermal insulation of the nebulizer and the connecting tubing from the external source heat. The elevated temperature of the ion source, ranging from 300 to 400 °C, serves a dual purpose: to offset the latent heat of vaporization during droplet desolvation and to change the solute into the gaseous state. When all parts of this straightforward interface are properly positioned and sized, each substance isolated by the nano-column transitions smoothly into the gas phase, the peak profile is accurately reproduced, and high-quality mass spectra are produced[13,14].

## Working Principle

The mixture of components to be separated is loaded on an autosampler, injected into an LC stream of mobile phase and separated on a column. The eluting fractions are then detected by a mass analyser (or commonly an in-line UV/DAD detector followed by a mass analyser) via ionisation in one of a number of techniques (EI, ESI, APCI, APPI, ESCI etc.). With respect to the LC setup, this can be HPLC or UPLC depending on the analytical method and there are obvious merits for each. There are a multitude of options on the characteristics of the mass analyser[15,16].





Figure 03: Principle of Liquid Chromatography-Mass Spectroscopy

## Instrumentation of Liquid Chromatography–Mass Spectrometry (LC–MS)

The LC–MS instrument consists of two major analytical units: a liquid chromatographic system (LC) for separating sample components and a mass spectrometer (MS) for ionizing, separating, and detecting the analytes according to their mass-to-charge ratio (\(m/z\)). These units are connected through an interface, usually an atmospheric-pressure ionization source.

## Major components and their functions

***Solvent reservoir:***

The solvent reservoir contains the mobile phase used to carry the sample through the chromatographic system. Common mobile-phase solvents include water, methanol, and acetonitrile, often combined with volatile additives such as formic acid or ammonium formate.

Function: Supplies the mobile phase required for chromatographic separation.

***Pump and degasser:***

The pump delivers the mobile phase through the system at a controlled flow rate and high pressure. The degasser removes

dissolved gases from the solvents to minimize bubble formation and fluctuations in detector response.

Function: Maintains a stable mobile-phase flow and reliable chromatographic performance.

***Sample injector or autosampler:***

The injector introduces a measured volume of sample into the mobile-phase stream. An autosampler automates sample injection and improves reproducibility during analysis of multiple samples.

Function: Introduces a precise quantity of sample into the LC system.

***Chromatographic column:***

The column contains the stationary phase responsible for separating the individual components of the sample. Reversed-phase C18 columns are widely used in pharmaceutical analysis.Separation occurs because different compounds interact differently with the stationary and mobile phases, resulting in different retention times.

Function: Separates the components of a mixture before mass spectrometric detection.

***Ionization source:***

The ionization source converts the compounds eluting from the LC column into gas-phase ions. Common ionization techniques include:

- Electrospray ionization (ESI): Suitable for polar, ionic, and many thermally labile compounds.
- Atmospheric-pressure chemical ionization (APCI): Often useful for relatively small, less polar compounds.
- Atmospheric-pressure photoionization (APPI): Useful for certain compounds that are difficult to ionize using ESI.

Function: Produces ions that can be analyzed by the mass spectrometer.

***Mass analyzer:***

The mass analyzer separates ions according to their mass-to-charge ratio (\(m/z\)). The choice of analyzer depends on the analytical purpose.

Table 01: Major Mass Analyzers and Their Applications in Mass Spectrometry

| Mass analyzer | Main application |
|---|---|
| Quadrupole | Routine analysis and quantitative measurements |
| Triple quadrupole | Sensitive and selective quantitative analysis using LC–MS/MS |
| Time-of-flight (TOF) | Accurate-mass measurement and compound identification |
| Ion trap | Ion storage and multistage fragmentation studies |
| Orbitrap | High-resolution and accurate-mass analysis |

Function: Differentiates ions according to their (m/z\) values.

***Detector***

The detector records the ions emerging from the mass analyzer and converts their signals into measurable data. The signal intensity is related to the abundance of the detected ions.

Function: Measures ion signals to support qualitative and quantitative analysis.

***Vacuum system***

The vacuum system maintains low pressure in the mass analyzer and associated regions where required. This minimizes collisions between ions and gas molecules, allowing efficient ion transmission and mass analysis.

Function: Provides the pressure conditions necessary for effective ion analysis.

***Computer and data-processing system***

Specialized software processes the signals obtained from the detector and displays the results as chromatograms and mass spectra.

- Chromatogram: Shows signal intensity against retention time.
- Mass spectrum: Shows ion abundance against mass-to-charge ratio.
- Quantitative analysis: Uses calibrated signal responses to determine analyte concentrations.

Function: Acquires, processes, stores, and interprets analytical data.

***Working sequence of LC–MS***

1. The sample is introduced into the liquid chromatographic system through an injector or autosampler.
2. The mobile phase transports the sample through the chromatographic column.
3. The column separates the sample components according to their physicochemical interactions.
4. The eluted compounds enter the ionization source and are converted into gas-phase ions.
5. The mass analyzer separates the ions according to their mass-to-charge ratio.
6. The detector measures the ion signals, and the computer generates chromatograms and mass spectra for interpretation [16-25].

## The use of LCMS in pharmaceutical process development

The use of LC-MS in a pharmaceutical setting is well established as analysts have always sought the best quality information from detectors and the outlay involved in set-up has significantly decreased. The seeding point for much of what is described below has been in the industrial QA/QC laboratories (due to the obvious qualitative and quantitative data produced) but also in academic collaborations. Recently, the application of LC-MS (in tandem with other qualitative and theoretical techniques) to process development has gained significant momentum. Outlined below are some recent developments in the use of LC-MS applied to process development change, a very significant area in pharmaceutical chemistry having achieved global investment in the past decade.

**Applications**

Liquid Chromatography–Mass Spectrometry (LC–MS) is widely used in pharmaceutical analysis, biomedical research, clinical diagnostics, environmental monitoring, and food safety. It combines the separation efficiency of liquid chromatography with the high sensitivity and selectivity of mass spectrometry, enabling the identification and quantification of compounds in complex samples.

*1. Pharmaceutical analysis*

- Identification and quantification of active pharmaceutical ingredients (APIs).

- Detection of impurities and degradation products.
- Analysis of drug formulations and stability studies.
- Development and validation of analytical methods.

*2. Pharmacokinetic and bioanalytical studies*

- Measurement of drugs and metabolites in plasma, serum, and urine.
- Evaluation of absorption, distribution, metabolism, and excretion (ADME).
- Bioavailability and bioequivalence studies.
- Therapeutic drug monitoring.

*3. Proteomics and metabolomics*

- Identification and characterization of proteins and peptides.
- Detection and quantification of metabolites.
- Biomarker discovery for disease diagnosis and prognosis.
- Investigation of biological pathways and disease mechanisms.

*4. Clinical diagnostics and toxicology*

- Detection of drugs and toxic substances in biological samples.
- Measurement of hormones, vitamins, and selected clinical biomarkers.
- Investigation of poisoning and drug overdose.
- Screening and confirmation of selected drugs of abuse.

*5. Food safety and quality control*

- Detection of pesticide residues in food products.
- Analysis of veterinary drug residues.
- Detection of mycotoxins and other contaminants.
- Identification of food additives and selected adulterants.

*6. Environmental analysis*

- Detection of pharmaceutical residues in water.
- Analysis of pesticides and industrial pollutants.
- Monitoring of emerging contaminants in wastewater.
- Assessment of chemical contamination in environmental samples.

*7. Drug discovery and development*

- Identification of drug candidates and metabolites.
- Characterization of synthesized compounds.
- Investigation of drug metabolism and metabolic pathways.
- Support for lead optimization and preclinical research.

*8. Forensic science*

- Detection of drugs and poisons in biological specimens.
- Analysis of toxicological evidence.
- Identification of selected unknown chemical substances.
- Support for forensic investigations [24-31].

Table 02: Specific Applications of LC–MS in Pharmaceutical Sciences

| Application | Purpose |
|---|---|
| Drug assay | Determines the concentration of an active pharmaceutical ingredient (API) in pharmaceutical formulations |
| Impurity profiling | Detects and characterizes impurities in drug substances and pharmaceutical products. |
| Stability studies | Identifies degradation products formed under different storage conditions. |
| Bioanalysis | Quantifies drugs and their metabolites in biological matrices such as plasma, serum, and urine. |
| Drug metabolism | Investigates metabolic transformations and metabolite formation in biological systems. |
| Herbal drug analysis | Identifies and quantifies selected phytochemicals and natural products in herbal formulations. |
| LC–MS/MS analysis | Enables sensitive and selective quantification using precursor-to-product ion transitions. |

## CONCLUSION

Recent developments in LC-MS technology have significantly expanded its application potential, opening new avenues in various scientific domains. This section explores the latest innovations and emerging trends in LC-MS, highlighting its transformative potential and future prospects. Current approaches to LC-MS correspondence are numerous and varied, yet share a host of drawbacks that must be considered during the design of next generation of algorithms. Because certain drawbacks are unavoidable with alignment approaches, correspondence-rather than warping functions-ought to be the focus of solutions. Methods with user-defined parameters need analytical and automatic solutions. All methods ought to have bounds on run time. Reference samples and fixed RT and m/z comparison windows are undesirable. Because there are already so many correspondence algorithms, we strongly suggest that any new algorithms abide by the above-mentioned criteria. LC–MS is an essential analytical technique because it provides high sensitivity, molecular-mass information, and the ability to analyze complex mixtures. Its applications extend from pharmaceutical quality control and drug development to clinical research, toxicology, food safety, and environmental monitoring. In pharmaceutical sciences, LC–MS/MS is particularly valuable for bioanalytical method development, pharmacokinetic studies, impurity characterization, and drug metabolism investigations.

## AUTHOR CONTRIBUTIONS

All authors contributed to equally.

## FUNDING

This research received no external funding.

## ACKNOWLEDGEMNET

Not Declared

## ETHICAL STATEMENT

Not Applicable

## REFERENCES

1. Ho CS, Lam CWK, Chan MHM, Cheung RCK, Law LK, Lit LCW, et al. Electrospray ionisation mass spectrometry: principles and clinical applications. Clin Biochem Rev. 2003;24:3-12.
2. Steinmann D, Ganzera M. Recent advances on HPLC/MS in medicinal plant analysis. Journal of Pharmaceutical and Biomedical Analysis. 2011; 55(4):744-757.
3. Sandeep B, Narender B. Stability-indicating liquid chromatographic method for the simultaneous estimation of colchicine and rosuvastatin in combination. Essential Chem. 2025; 2(1):1-11.
4. Srinidhi M, Basha MM, Kumar VR, Kumar JR. Stability indicating RP-HPLC method development and validation for the estimation of sumatriptan in bulk and pharmaceutical dosage form. Journal of Applied Pharmaceutical Science. 2016;6(6):020-025.
5. Bolla S, BoggulaN. Quantitative determination of Nebivolol Hydrochloride and Rosuvastatin Calcium in laboratory prepared mixture by robust, high-performance liquid chromatographic method. Discov. Chem. 2025; 2:61.
6. Nagula S, Thota S, Kumar VR. Development and validation of a stability indicating RP-HPLC method for the determination of trimetazidine dihydrochloride. Der Pharma Chemica. 2014;6(1):149-54.
7. Narasimha K, Bhuvanachandra P, Narender B, Praveen KK, Daniel K, Vamshikrishna G, Kiran G. A HPLC-MS/MS method for the determination of Nadolol in rat plasma: Development, validation, and application to pharmacokinetic study. European Journal of Mass Spectrometry. 2023; 29(3):170-180.
8. Pokkula S, Thota S, Kumar VR, Nagabandi VK. Development and validation of RP-HPLC method for the determination of methylphenidate hydrochloride in API. International Journal of PharmTech Research. 2014;6:462-7.
9. Sushma A,Shravan KK, Narender B, Suchitra D. Development and Validation of Novel Bioanalytical Method for Simultaneous Estimation of Simvastatin, Ramipril, Atenolol, Hydrochlorothiazide and Aspirin in Polycap™ Capsule Using LC-MS/MS. International Journal of Zoological Investigations. 2024; 10(2):652-663.
10. Alanazi S. Recent Advances in Liquid Chromatography-Mass Spectrometry (LC-MS) Applications in Biological and Applied Sciences. Anal Sci Adv. 2025;6(1):e70024.
11. Narender B, Shanmuga Pandiyan P. Development and Validation of RP-HPLC Method for the Simultaneous Estimation of Dapagliflozin and Saxagliptin in Bulk and Pharmaceutical Dosage Forms. Int J Pharm Sci & Res. 2021; 12(1):314-320.
12. Gumulapuram SH, Thota SR, Kumar VR, Nagabandi VK. Stability indicating RP-HPLC method development and validation for the simultaneous estimation of eprosartan mesylate and hydrochlorothiazide in bulk and tablet dosage form. Int. J. Pharm. Pharm. Sci. 2014;6(2):446-451.
13. Narender B, Shanmuga Pandiyan P. Development and Validation of Opicapone in Human Plasma Using Tolcapone as Internal Standard by Liquid Chromatography-Tandem Mass Spectrometry. Journal of Emerging Technologies and Innovative Research. 2019; 6(1):1075-1089.
14. Beccaria M, Cabooter D.Current Developments in LC-MS for Pharmaceutical Analysis. Analyst. 2020; 145:1129–1157.
15. Sravani T, Thota S, Venisetty R, Venumadhav N. RP-HPLC analysis of acebrophylline in API and capsule dosage form. Res J Pharm. 2014;5:480-486.
16. Vamshikrishna G, Bhuvanachandra P, Heeralal B, Naveen P, Narender B. A Validated UPLC-MS/MS Method for the Quantification of Cyclophosphamide in Human Plasma: Application to Therapeutic Drug Monitoring in Cancer Patients. *Annals* of the Romanian Society for Cell Biology. 2021; 25(6):15063-15080.
17. Meghana M, Thota S, Venisetty RK. Development and validation of stability-indicating RP-HPLC method for the estimation of Agomelatine in API. Research Journal of Pharmaceutical, Biological and Chemical Sciences. 2014;5(1):621-628.
18. Revathi B, Jithendar Reddy M, Narender B, Rajeshwar V, Suchitra D. Development of Expeditious RP-HPLC Method for Simultaneous Estimation of Bilastine and Montelukast in Fixed Dosage Combination. Journal of Applied Pharmaceutical Research. 2025; 13(6):113-119.
19. Pullen F.The Fascinating History of the Development of LC-MS; A Personal Perspective. Chromatography Today. 2010; 3:4–6.
20. Boggula N, Bhadru B, More K. RP-HPLC Method Validation for Levomilnacipran Estimation in Bulk and Formulation. International Journal of Pharmaceutical Quality Assurance. 2023; 14(4):900-903.
21. Díaz-Cruz MS, Barceló D.Recent Advances in LC-MS Residue Analysis of Veterinary Medicines in the Terrestrial Environment. Trends in Analytical Chemistry. 2007; 26:637–646.
22. Akiful Haque M, Shanthi Priya DK, Mohanty D, Vasudha B, Narender B. Analytical Method Development and Validation of Amlodipine in Human Plasma by Using Liquid Chromatography–Mass Spectrometry/Mass Spectrometry. *Asian J Pharm Clin Res. 2018; 11(7):393-397.*
23. Ramu K, Aleti P, Venisetty RK. Analytical Method Development and Validation of Simultaneous Estimation of Ezetamibe and Rosuvastatin in Tablet Dosage Form By RP-HPLC. Res. Art. Pharm. Sci. 2013;3(4):343-353.
24. Sriveni T, Naveen V, Sai Rupa V, Renuka A, Sunil P, Akiful Haque M, Vasudha B, Narender B. Development and Validation of Dolutegravir in Bulk and Formulation: An Anti-Retroviral Drug Using UV-Spectroscopy. International

Journal of Pharmaceutical Quality Assurance. 2021; 12(1):57-60.

25. Kudikala S, Malladi SR, Thota S, Kumar VR. RP-HPLC method for the estimation gemcitabine in API and parenteral dosage form. Journal of Scientific Research in Pharmacy. 2014;3:16-18.
26. Narender B, Gowri Manoja M, Shanthi Priya DK, Vasudha B, Himabindu P. Method Development and Validation for the Simultaneous Estimation of Rosuvastatin and Amlodipine in Bulk and its Formulation using Reverse-Phase High-Performance Liquid Chromatography. Asian Journal of Pharmaceutics. 2018; 12(4):S1304-S1309.
27. Swapna B, Kiran G, Vasudha B, Kumar JR. Stability indicating RP-HPLC method for simultaneous estimation of betamethasone dipropionate and calcipotriene in bulk and pharmaceutical dosage form. Biointerface Research in Applied Chemistry. 2018;8(1):3089-94.
28. Sowmya Y, Aleti P, Venisetty R. Development and validation of RP-HPLC method for the simultaneous estimation of dutasteride and tamsulosin in tablet dosage form. International Journal of Pharmacy and Biological Sciences. 2013;3(4):301-316.
29. Keshetty S, Venisetty RK, Molmoori V, Ciddi V. Determination of valdecoxib in serum using a HPLC-diode array detector and its application in a pharmacokinetic study. Die Pharmazie-An International Journal of Pharmaceutical Sciences. 2006;61(3):245-246.
30. Battineni JK, Mamidisetti YD, Kuchukuntla M, Amaravadi D, Ravula SR, Jadi RK. Mass Spectrometry Based Multi-Omics Integration: Analytical Methods and Biological Applications. Rapid Communications in Mass Spectrometry. 2026;40(15):e70097.
31. Vekateswarlu K, Venisetty RK, Yellu NR, Keshetty S, Pail MG. Development of HPTLC-UV absorption Densitometry and Sertraline in combination and its application in evaluation of marketed preparation. Journal of Chromatographic Science. 2007;45(8):537–539.